

# ~ Calibration Certificate ~

Per ISO 16063-21

Model Number: 352C42

Serial Number: 69907

Description: ICP® Accelerometer


Method: Back-to-Back Comparison Calibration

Manufacturer: PCB

## Calibration Data

Sensitivity @ 100.0 Hz	104.8 mV/g	Output Bias	13.9 VDC
	(10.68 mV/m/s <sup>2</sup> )	Transverse Sensitivity	0.9 %
Discharge Time Constant	0.7 seconds	Resonant Frequency	31.2 kHz

## Sensitivity Plot



## Data Points

Frequency (Hz)	Dev. (%)	Frequency (Hz)	Dev. (%)	Frequency (Hz)	Dev. (%)
10.0	2.7	300.0	-1.0	7000.0	0.1
15.0	2.2	500.0	-1.4	10000.0	4.7
30.0	1.3	1000.0	-2.0		
50.0	0.7	3000.0	-2.5		
REF. FREQ.	0.0	5000.0	-1.5		

Mounting Surface: Stainless Steel    Fastener: Cyanoacrylate Adhesive  
Acceleration Level (rms): 10.0 g (98.1 m/s<sup>2</sup>)

Fixture Orientation: Vertical

<sup>1</sup>The acceleration level may be limited by shaker displacement at low frequencies. If the listed level cannot be obtained, the calibration system uses the following formula to set the vibration amplitude: Acceleration Level (g) = 0.010 x (freq)<sup>2</sup>.

<sup>2</sup>The gravitational constant used for calculations by the calibration system is; 1 g = 9.80665 m/s<sup>2</sup>.

## Condition of Unit

As Found: n/a

As Left: New Unit, In Tolerance

## Notes

1. Calibration is NIST Traceable thru Project 822/271196 and PTB Traceable thru Project 5399.
2. This certificate shall not be reproduced, except in full, without written approval from PCB Piezotronics, Inc.
3. Calibration is performed in compliance with ISO 9001, ISO 10012-1, ANSI/NCSL Z540-1-1994 and ISO 17025.
4. See Manufacturer's Specification Sheet for a detailed listing of performance specifications.
5. Measurement uncertainty (95% confidence level with coverage factor of 2) for frequency ranges tested during calibration are as follows: 5-9 Hz; +/- 2.0%, 10-99 Hz; +/- 1.5%, 100-1999 Hz; +/- 1.0%, 2-10 kHz; +/- 2.5%.

Technician: Dwight Beverly

Date: 06/07/06



**PCB PIEZOTRONICS**  
VIBRATION DIVISION

3425 Walden Avenue    Depew, NY 14043

TEL: 888-684-0013    FAX: 716-685-3886    www.pcb.com

# ~ Calibration Certificate ~

Per ISO 16063-21

Model Number: 352C22

Serial Number: 73349

Description: ICP® Accelerometer


Method: Back-to-Back Comparison Calibration

Manufacturer: PCB

### Calibration Data

Sensitivity @ 100.0 Hz	10.12 mV/g	Output Bias	10.2 VDC
	(1.032 mV/m/s <sup>2</sup> )	Transverse Sensitivity	1.9 %
Discharge Time Constant	2.6 seconds	Resonant Frequency	86.5 kHz

### Sensitivity Plot



### Data Points

Frequency (Hz)	Dev. (%)	Frequency (Hz)	Dev. (%)	Frequency (Hz)	Dev. (%)
10.0	-0.6	300.0	0.2	7000.0	1.4
15.0	-0.6	500.0	0.2	10000.0	2.5
30.0	-0.2	1000.0	0.3		
50.0	-0.2	3000.0	0.4		
REF. FREQ.	0.0	5000.0	0.8		

Mounting Surface: Tungsten Adapter    Fastener: Cyanoacrylate Adhesive  
 Acceleration Level (rms)<sup>1</sup>: 10.0 g (98.1 m/s<sup>2</sup>)

Fixture Orientation: Vertical

<sup>1</sup>The acceleration level may be limited by shaker displacement at low frequencies. If the listed level cannot be obtained, the calibration system uses the following formula to set the vibration amplitude: Acceleration Level (g) = 0.010 x (freq)<sup>2</sup>.  
<sup>2</sup>The gravitational constant used for calculations by the calibration system is: 1 g = 9.80665 m/s<sup>2</sup>.

### Condition of Unit

As Found: n/a  
 As Left: New Unit, In Tolerance

### Notes

1. Calibration is NIST Traceable thru Project 822/271196 and PTB Traceable thru Project 5399.
2. This certificate shall not be reproduced, except in full, without written approval from PCB Piezotronics, Inc.
3. Calibration is performed in compliance with ISO 9001, ISO 10012-1, ANSI/NCSL Z540-1-1994 and ISO 17025.
4. See Manufacturer's Specification Sheet for a detailed listing of performance specifications.
5. Measurement uncertainty (95% confidence level with coverage factor of 2) for frequency ranges tested during calibration are as follows: 5-9 Hz; +/- 2.0%, 10-99 Hz; +/- 1.5%, 100-1999 Hz; +/- 1.0%, 2-10 kHz; +/- 2.5%.

Technician: Kevin Moseley *KM*      Date: 05/16/06



Headquarters: 3425 Walden Avenue, Depew, NY 14043  
 Manufacturing and Calibration Facility: 10869 Highway 903, Halifax, NC 27839  
 TEL: 888-684-0013    FAX: 716-685-3886    www.pcb.com

# ~ Calibration Certificate ~

Per ISO 16063-21

Model Number: 352A73

Serial Number: 160588

Description: ICP® Accelerometer


Manufacturer: PCB

Method: Back-to-Back Comparison AT401-3

### Calibration Data

Sensitivity @ 100 Hz	4.44 mV/g (0.453 mV/m/s <sup>2</sup> )	Output Bias	9.9 VDC
		Transverse Sensitivity	2.7 %
Discharge Time Constant	0.5 seconds	Resonant Frequency	107.8 kHz

### Sensitivity Plot



### Data Points

Frequency (Hz)	Dev. (%)	Frequency (Hz)	Dev. (%)	Frequency (Hz)	Dev. (%)
10	-0.5	300	0.6	7000	2.0
15	0.0	500	0.7	10000	2.4
30	0.3	1000	1.0		
50	0.4	3000	1.5		
REF. FREQ.	0.0	5000	1.8		

Mounting Surface: Tungsten Adapter Fastener: Adhesive Fixture Orientation: Vertical

Acceleration Level (pk): 10.0 g (98.1 m/s<sup>2</sup>)

\*The acceleration level may be limited by shaker displacement at low frequencies. If the listed level cannot be obtained, the calibration system uses the following formula to set the vibration amplitude: Acceleration Level (g) = 0.008 x (freq)<sup>2</sup>. \*The gravitational constant used for calculations by the calibration system is: 1 g = 9.80665 m/s<sup>2</sup>.

### Condition of Unit

As Found: n/a

As Left: New Unit, In Tolerance

### Notes

1. Calibration is NIST Traceable thru Project 683/283498 and PTB Traceable thru Project 10065.
2. This certificate shall not be reproduced, except in full, without written approval from PCB Piezotronics, Inc.
3. Calibration is performed in compliance with ISO 9001, ISO 10012-1, ANSI Z540.3 and ISO 17025.
4. See Manufacturer's Specification Sheet for a detailed listing of performance specifications.
5. Measurement uncertainty (95% confidence level with coverage factor of 2) for frequency ranges tested during calibration are as follows: 5-9 Hz; +/- 2.0%, 10-99 Hz; +/- 1.5%, 100-1999 Hz; +/- 1.0%, 2-10 kHz; +/- 2.5%.

Technician: Donald Whalen *DW* Date: 6/6/2013



CALIBRATION CERT #1862.02

PAGE 1 of 1



VIBRATION DIVISION

Headquarters: 3425 Walden Avenue, Depew, NY 14043

Calibration Performed at: 10869 Highway 903, Halifax, NC 27839

TEL: 888-684-0013 FAX: 716-685-3886 www.pcb.com

CAL48-3453364761.646+0

