

Certificate of Calibration and Conformance

Certificate Number 2010-129256

Instrument Model CAL200, Serial Number 7672, was calibrated on 30APR2010. The instrument meets factory specifications per Procedure D0001.8190.

New Instrument

Date Calibrated: 30APR2010

Calibration due:

Calibration Standards Used

MANUFACTURER	MODEL	SERIAL NUMBER	INTERVAL	CAL. DUE	TRACEABILITY NO.
Hewlett Packard	34401A	US36033460	12 Months	16JUN2010	4382218
Hewlett Packard	34401A	3146A10352	12 Months	13JUL2010	4413817
Larson Davis	PRM915	0112	12 Months	09SEP2010	2009-121809
Larson Davis	PRM902	0480	12 Months	09SEP2010	2009-121820
Larson Davis	MTS1000/2201	0111	12 Months	09SEP2010	SM090909-1
Larson Davis	2559	2504	12 Months	29SEP2010	16910-1
PCB	1502B02FJ15PSIA	1342	12 Months	23NOV2010	3341845067
Larson Davis	2900	0661	12 Months	02APR2011	2010-128279

Reference Standards are traceable to the National Institute of Standards and Technology (NIST)

Calibration Environmental Conditions

Environmental test conditions as shown on calibration report.


Affirmations

This Certificate attests that this instrument has been calibrated under the stated conditions with Measurement and Test Equipment (M&TE) Standards traceable to the U.S. National Institute of Standards and Technology (NIST). All of the Measurement Standards have been calibrated to their manufacturers' specified accuracy / uncertainty. Evidence of traceability and accuracy is on file at Provo Engineering & Manufacturing Center. An acceptable accuracy ratio between the Standard(s) and the item calibrated has been maintained. This instrument meets or exceeds the manufacturer's published specification unless noted.

This calibration complies with the requirements of ISO 17025 and ANSI Z540. The collective uncertainty of the Measurement Standard used does not exceed 25% of the applicable tolerance for each characteristic calibrated unless otherwise noted.

The results documented in this certificate relate only to the item(s) calibrated or tested. A one year calibration is recommended, however calibration interval assignment and adjustment are the responsibility of the end user. This certificate may not be reproduced, except in full, without the written approval of the issuer.

Signed:



Technician: Scott Montgomery

Larson Davis CAL200 Acoustic Calibrator, SN: 7672
Certificate of Measured Output

Performance at Reference Conditions

Nominal Level (dB SPL):	94	114
Measured Level (dB SPL):	94.02	114.00
Expanded Uncertainty (dB):	0.154	0.149
Level Error Limit (dB):	±0.35	±0.35
Nominal Frequency (Hz):	1000	1000
Measured Frequency (Hz):	1000.0	1000.0
Expanded Uncertainty (Hz):	0.2	0.2
Frequency Error Limit (Hz):	±10.0	±10.0
Measured Distortion (%):	0.57	0.34
Expanded Uncertainty (%):	0.25	0.25
Distortion Limit (%):	2.0	2.0

The data is acquired by the insert voltage calibration method using the reference microphone's open circuit sensitivity.

Environmental Conditions

Temperature (°C):	22	22
Relative Humidity (%):	31	32
Static Pressure (kPa):	101.2	101.1

Reference Microphone

Model: Larson Davis 2559
Serial Number: 2504
Open Circuit Sensitivity: 11.444 mV/Pascal
Uncertainty: 0.130 dB

Influence of Static Pressure

Nominal Level (dB SPL):		114		
Nominal Pressure (kPa)	Pressure (kPa)	Level Change (dB)	Frequency Change (Hz)	Distortion (%)
108.0	107.9	-0.03	0.00	0.33
101.3	101.2	0.00	0.00	0.33
92.0	92.0	0.04	0.00	0.34
83.0	83.0	0.04	0.00	0.36
74.0	73.9	-0.00	0.00	0.38
65.0	65.1	-0.12	0.00	0.41
Expanded Uncertainty:	1.0	0.04	0.20	0.25
Limit:		±0.30	±10.0	2.0

Reference microphone corrections applied.

Environmental Conditions

Temperature (°C):	22
Relative Humidity (%):	39

Reference Microphone

Model: Larson Davis 2559
Serial Number: 2506

Static pressure was measured with a calibrated Motorola pressure sensor MPX2100AF.
Temperature and humidity was measured with a calibrated Fluke 1620A sensor.
Expanded uncertainty of environmental measurements: 0.3 °C, 3 %RH, 1.0 kPa
Uncertainty values are given at 95% confidence level (k = 2).


A Sound Level Meter can be calibrated to a level (L) defined as: L = measured level + pressure sensitivity
or if a Sound Level Meter is calibrated using the nominal level, the adjustments to data (X) are defined as:
X = measured level - nominal level - pressure sensitivity



Model CAL200 Relative SPL vs. Temperature
Larson Davis Model CAL200 Serial Number: 7672

Model CAL200 Relative SPL vs. Temperature at 50% RH.
A 2259 Mic (SN: 2993) with a PRM901 Preamp (SN: 0208), station 17 was used to check the levels.

Test Date: 20 Apr 2010 16:56:50



0.1dB expanded uncertainty at ~95% confidence level (k=2)

Test performed at: Larson Davis, a division of PCB Piezotronics, Inc.
1681 West 820 North, Provo, Utah 84601


Larson Davis, a division of PCB Piezotronics, Inc.
Tel: 716 684-0001 www.LarsonDavis.com



Model CAL200 Relative Frequency vs. Temperature
Larson Davis Model CAL200 Serial Number: 7672

Model CAL200 Relative Frequency vs. Temperature at 50% RH.
A 2259 Mic (SN: 2993) with a PRM901 Preamp (SN: 0208), station 17 was used to check the levels.

Test Date: 20 Apr 2010 16:56:50



1.0 Hz expanded uncertainty at ~95% confidence level (k=2)

Test performed at: Larson Davis, a division of PCB Piezotronics, Inc.
1681 West 820 North, Provo, Utah 84601

Larson Davis, a division of PCB Piezotronics, Inc.
Tel: 716 684-0001 www.LarsonDavis.com