

~Certificate of Calibration~

Model Number: 483C15

PCB Control #: QC214/QC184/QC198/CA514

Serial Number: 133

Calibration Date: 01/02/2009

Description: Signal Conditioner

Recalibration Date:

Test Procedure: AT104-29

Calibration Technician: Chris Vega 301

Temperature: 72° F

Relative Humidity: 36%

Channel	Volts	Current (mA)	Gain X1	Gain X10	Gain X100
1	26.0	4.0	0.999	9.99	99.8
2	26.0	4.0	0.999	10.00	99.8
3	26.0	4.0	0.999	10.01	99.8
4	26.0	4.0	0.999	10.00	99.8
5	25.8	4.0	0.999	10.01	99.9
6	25.8	4.0	0.999	9.99	99.8
7	25.8	4.0	0.999	10.00	99.9
8	25.8	4.0	0.999	10.00	99.9

As Received: New Unit.

As Left: In tolerance.

Special Notes:

This document certifies that the equipment referenced below meets published specifications. The calibration procedure is in compliance with ISO 10012-1, and former MIL-STD-45662A and is traceable to NIST. Measurement uncertainty (95% confidence level w/coverage factor of 2) for scale factors is +/- 0.2%.

This certificate may not be reproduced, except in full, without written approval of PCB Piezotronics, Inc.



PCB PIEZOTRONICS™

3425 Walden Avenue Depew, New York, USA 14043-2495

For any questions concerning this certificate, please call PCB at (716) 684-0001 and ask for an application engineer.